

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Accredited to  
ISO/IEC 17025:2005

### Davidson and Hardy (Laboratory Supplies) Limited

Issue No: 006 Issue date: 27 April 2017

453 – 455 Antrim Road  
Belfast  
BT15 3BL

Contact: Mr J Featherstone  
Tel: +44 (0)7501 255405  
Fax: +44 (0)28 9037 0115  
E-Mail: joe@dhlab.com  
Website: www.dhlab.com

Calibration performed by the Organisation at the locations specified below

Location details	Activity	Location code
<p>Customer Premises – Any</p> <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p><b>Local contact</b> Mr J Featherstone (as above)</p> <p>Electrical Temperature Rotational speed</p>	<p>S</p>



7783  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Davidson and Hardy (Laboratory Supplies), Limited**  
Issue No: 006 Issue date: 27 April 2017

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
<b>ELECTRICAL</b> Time Interval	10 s to 10 hours	1.0 s		S
<b>TEMPERATURE</b> Temperature controlled autoclaves, media preparators and sterilizers (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	- 25 °C to 140 °C	0.42 °C	Single and multipoint monitoring probes. Time dependent temperature profiling.	S
Incubators	- 25 °C to 90 °C	0.42 °C		S
Liquid baths	- 25 °C to 100 °C	0.42 °C		S
Environmental chambers	- 25 °C to 40 °C	0.42 °C		S
Digital Indicators with Probes	-25 °C to 140 °C	0.11 °C		S
<b>ROTATIONAL SPEED</b> Centrifuge	50 rpm to 60000 rpm	3 rpm		S
END				